

CERTIFICATE OF ANALYSIS

CHROMOGENIX

PRODUCT:	S-2302, 25 mg Substrate	PRODUCT #:	00082034039
EXP. DATE:	2018-03	LOT #:	N1143456
VENDOR:	INSTRUMENTATION LABORATORY	ASSAYED PER QCM:	820340

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION
	MIN	MAX			PASS/FAIL
SUBSTRATE CONCENTRATION	23.7	26.3	mg/vial	24.5	PASS
FREE p-NITROANILINE	0	0.5	mol %	0.0	PASS

DESCRIPTION	LOT NUMBER	EXP. DATE	VOLUME
S-2302	N1143456	2018-03	3 mL

PREPARED BY: M. Oettinger *M. Oettinger* DATE: 12/12/2014 ¹²⁻¹²⁻¹⁴

APPROVAL SIGNATURE: D. Tencza *D. Tencza* DATE: 12/12/2014 ^{12/12/14}

Form QCM00820340-2 Rev. 03/09