

CERTIFICATE OF ANALYSIS

CHROMOGENIX

PRODUCT:	S-2302, 25 mg Substrate	PRODUCT #:	00082034039
EXP. DATE:	2019-03	LOT #:	N1159415
VENDOR:	INSTRUMENTATION LABORATORY	ASSAYED PER QCM:	820340

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION
	MIN	MAX			
SUBSTRATE CONCENTRATION	23.7	26.3	mg/vial	24.4	PASS
FREE p-NITROANILINE	0	0.5	mol %	0.0	PASS

DESCRIPTION	LOT NUMBER	EXP. DATE	VOLUME
S-2302	N1159415	2019-03	3 mL

PREPARED BY: Maria Battenger DATE: 11-24-15

APPROVAL SIGNATURE: Debrae Woz DATE: 11-25-15

Form QCM00820340-2 Rev. 03/09