

CERTIFICATE OF ANALYSIS

CHROMOGENIX

PRODUCT:	S-2288, 25 mg Substrate	PRODUCT #:	82085239
EXP. DATE:	2015-01	LOT #:	N0128973
VENDOR:	INSTRUMENTATION LABORATORY	ASSAYED PER QCM:	820852

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION
	MIN	MAX			
SUBSTRATE CONCENTRATION	23.7	26.3	mg/vial	25.2	PASS
FREE p-NITROANILINE	0	0.5	mol %	0	PASS

DESCRIPTION	LOT NUMBER	EXP. DATE	VOLUME
S-2288	N0128973	2015-01	2 ml

PREPARED BY: M. Oettinger DATE: 2/17/2012

APPROVAL SIGNATURE: D. Tencza DATE: 2/17/2012