

CERTIFICATE OF ANALYSIS

CHROMOGENIX

PRODUCT: S-2765, 25 mg Substrate	PRODUCT #: 82141339
EXP. DATE: 2015-02	LOT #: N0229028
VENDOR: INSTRUMENTATION LABORATORY	ASSAYED PER QCM: 821413

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION
	MIN	MAX			
SUBSTRATE CONCENTRATION	23.7	26.3	mg/vial	25.0	PASS
FREE p-NITROANILINE	0	0.5	mol %	0	PASS

DESCRIPTION	LOT NUMBER	EXP. DATE	VOLUME
S-2765	N0229028	2015-02	3 ml

PREPARED BY: M. Oettinger *M. Oettinger* DATE: 2/21/2012 *2-21-12*

APPROVAL SIGNATURE: D. Tencza *J. Dencza* DATE: 2/21/2012 *2/21/12*