

CERTIFICATE OF ANALYSIS

CHROMOGENIX

| | |
|------------------------------------|-------------------------|
| PRODUCT: S-2765, 25 mg Substrate | PRODUCT #: 82141339 |
| EXP. DATE: 2016-03 | LOT #: N0333612 |
| VENDOR: INSTRUMENTATION LABORATORY | ASSAYED PER QCM: 821413 |

| TEST DESCRIPTION | SPECIFICATIONS | | UNITS | RESULTS | DISPOSITION PASS/FAIL |
|-------------------------|----------------|------|---------|---------|--------------------------|
| | MIN | MAX | | | |
| SUBSTRATE CONCENTRATION | 23.7 | 26.3 | mg/vial | 25.1 | Pass |
| FREE p-NITROANILINE | 0 | 0.5 | mol % | 0 | Pass |

| DESCRIPTION | LOT NUMBER | EXP. DATE | VOLUME |
|-------------|------------|-----------|--------|
| S-2765 | N0333612 | 2016-03 | 3 ml |

PREPARED BY: M. Oettinger *M. Oettinger* DATE: 4/4/2013 *4-4-13*

APPROVAL SIGNATURE: D. Tenca *D. Tenca* DATE: 4/4/2013 *4/5/13*