CERTIFICATE OF ANALYSIS

CHROMOGENIX

PRODUCT:	S-2765, 25 mg Substrate	PRODUCT # :	82141339
EXP. DATE:	2017-11	LOT # :	N1143454
VENDOR:	INSTRUMENTATION LABORATORY	ASSAYED PER QCM:	821413

TEST DESCRIPTION	SPECIFICATIONS			RESULTS	DISPOSITION
	MIN	MAX	UNITS		PASS/FAIL
SUBSTRATE CONCENTRATION	23.7	26.3	mg/vial	25.4	PASS
FREE p-NITROANILINE	0	0.5	mol %	0	PASS

DESCRIPTION	LOT NUMBER	EXP.DATE	VOLUME
S-2765	N1143454	2017-11	3 ml

PREPARED BY: M. Oettinger M Oettinger DATE: 12/15/2014 12-15-14

12/15/14

APPROVAL SIGNATURE: D. Tenoza D Tenoza DATE: 12/15/2014

526 Route 303, Orangeburg, New York 10962, Telephone 845-365-8000