



Certificate of Analysis

Factor XIa Control

REF	Catalogue number	1188
LOT	Batch code	HHN0418
	Use by	2021 06
	Upper limit of temperature	-20°C


Activity after reconstitution with **4.0 mL** water:

Acceptance range
8.9 - 13.3 mIU/mL

The above Factor XIa activity apply only to **Lot#HHN0418**.

The calibration was performed using the Rox Factor XIa kit (Ref#110050). Factor XIa Calibrator (Ref#1199), potency assigned vs the 1st International standard for Activated Blood Coagulation Factor XI (FXIa), Human, NIBSC code 13/100, was used as calibrator.

The **Lot#HHN0418** has been approved for release after analyses according to established protocols with results demonstrating fulfillment of product specifications. All data concerning materials used, production data and release testing are documented and traceable under the individual Lot# which is stated on each package.



Pia Bryngelhed
Quality Manager, Rossix AB

May 18, 2016

Date



Rossix AB
Taljegardsgatan 3B
SE-431 53 Molndal
Sweden

Phone: +46 31 706 8965
Fax: +46 31 706 8966
info@rossix.com
www.rossix.com