

# CERTIFICATE OF ANALYSIS





DiaPharma Group, Inc.  
8948 Beckett Road, West Chester, OH 45069

<b>Product:</b>	DiaPharma Factor X kit
<b>Item Number:</b>	DPGFX
<b>Vendors:</b>	R <sup>2</sup> Diagnostics, Chromogenix
<b>Lot #:</b>	DPGFX045
<b>Exp. Date:</b>	02/2018

Kit Components:			
Name	Lot Number	Exp. Date	Volume (1 vial each)
FXa Chromogenic Substrate	N0160372	01/2019	25 mg
Russell's Viper Venom	DR-963P	02/2018	<1 mg
CaCl <sub>2</sub>	DC-963P	02/2018	20 ml
Tris Buffer	DB-963P	02/2018	100 ml

Test Description	Specifications		Results		Disposition
	Min	Max	Units	Value	Pass/Fail
Slope	1.0e-3 to 4.0e-3		NA	3.240e-03	Pass
Intercept	-.09 to .09		NA	1.960e-03	Pass
Regression Coefficient	R > 0.990		NA	0.999	Pass

Prepared By:  Date: 08/31/2016  
Approved By:  Date: 9/13/16