

**CERTIFICATE OF ANALYSIS**



DiaPharma Group, Inc.  
8948 Beckett Road, West Chester, OH 45069

<b>Product:</b>	DiaPharma Factor X kit
<b>Item Number:</b>	DPGFX
<b>Vendors:</b>	R <sup>2</sup> Diagnostics, Chromogenix
<b>Lot #:</b>	DPGFX046
<b>Exp. Date:</b>	02/2018

<b>Kit Components:</b>			
<b>Name</b>	<b>Lot Number</b>	<b>Exp. Date</b>	<b>Volume (1 vial each)</b>
FXa Chromogenic Substrate	N0462361	04/2019	25 mg
Russell's Viper Venom	DR-971	02/2018	<1 mg
CaCl <sub>2</sub>	DC-963P	02/2018	20 ml
Tris Buffer	DB-963P	02/2018	100 ml

<b>Test Description</b>	<b>Specifications</b>		<b>Results</b>		<b>Disposition</b>
	Min	Max	Units	Value	
Slope	1.0e-3	4.0e-3	NA	3.123e-03	Pass
Intercept	-0.09	0.09	NA	5.712e-03	Pass
Regression Coefficient	R > 0.990		NA	0.999	Pass

Prepared By: *[Signature]* Date: 08/16/2016

Approved By: *[Signature]* Date: 9/13/2016