



Certificate of Analysis

REAADS
Monoclonal Free
Protein S Test Kit

Kit Lot #: MFPS-181
Plate Lot #: FSW-063
Product #: 051-001
Expiration Date: 2018-04-03

Quality Control Results

Testing Date: 2016-10-26
Manufacture Date: 2016-10-03

| Absorbance | Value | Specification | Result by Assay |
|--------------------------------|----------------|----------------------|------------------------|
| Reference Plasma 100% OD | 86 % of Normal | ≥ 1.0 | 1.350 |
| Reagent Blank | | ≤ 0.250 | 0.091 |
| Reference Curve R ² | | ≥ 0.95 | 0.99791 |

| Control Response | Label Range | Result by Assay |
|-------------------------|--------------------|------------------------|
| Normal Control | 79 – 119 % | 96.4 % |
| Deficient Control | 24 – 55 % | 41.6 % |

This signed Certificate of Analysis affirms that the product has met the requirements of the relevant specifications as set by Corgenix, Inc.

This product is warranted to perform as described in the package insert. Corgenix, Inc. disclaims any implied warranty of merchantability or fitness for a particular use, and in no event shall Corgenix, Inc. be liable for consequential damage.

QA Approval By: Christin J. Bergano

Signed:  2016-11-15

Title: Quality Assurance Associate

Corgenix, Inc.
11575 Main Street, Suite 400
Broomfield, CO 80020
Phone: (303) 457-4345
Fax: (303) 457-4519