

CERTIFICATE OF ANALYSIS

CHROMOGENIX

PRODUCT:	Antithrombin 1 x 25 IU	PRODUCT # :	81079639
EXP. DATE:	2018-12	LOT # :	N1267001
VENDOR:	INSTRUMENTATION LABORATORY	ASSAYED PER QCM:	654171

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION
	MIN	MAX			
Antithrombin Activity	25	30	IU/vial	30	PASS
Assay Slope @ 37°C	-0.0024	-0.0034		-0.0029	PASS
Intercept @ 37°C	0.40	0.60		0.44	PASS
r ²	0.990	--		0.999	PASS

KIT COMPONENTS :			
NAME	LOT NUMBER	EXP. DATE	VOLUME
Antithrombin, 25 IU	N1267001	2018-12	1 ml

PREPARED BY: *M. Oettinger*

DATE: 1/12/2017

APPROVAL SIGNATURE: *D. Tencza*

DATE: 1/12/2017