

CERTIFICATE OF ANALYSIS

CHROMOGENIX

PRODUCT:	COATEST APC Resistance Control Plasma Level 1	PRODUCT #:	82265063
EXP. DATE:	2018-12	LOT #:	N1266700
VENDOR:	INSTRUMENTATION LABORATORY	ASSAYED PER QCM:	C00423

TEST DESCRIPTION	SPECIFICATIONS			RESULTS	DISPOSITION	
	MIN	MAX	UNITS			
Analysis with COATEST APC R:						
APT Time	CaCl ₂	28	34	seconds	30	PASS
	APC/CaCl ₂	90	--	seconds	117	PASS
APC Ratio		3.5	4.4	--	4.0	PASS

Analysis with COATEST APC R V:						
APT Time	CaCl ₂	--	38	seconds	35	PASS
	APC/CaCl ₂	75	--	seconds	109	PASS
APC Ratio		2.7	3.2	--	3.1	PASS

Kit Components:

DESCRIPTION	LOT NUMBER	EXP. DATE	VOLUME
Control Plasma Level 1	N1266700	2018-12	1 ml (x 5)

We hereby certify that the above product(s):

- have been manufactured and tested according to the requirements of their Manufacturing and Quality Control procedures and have been found to meet all conditions set forth with these documents;
- meet all applicable requirements of the 98/79/EC Directive according to which they are CE marked.

PREPARED BY: MOETTINGER M. Moettinger DATE: 12/12/2016 12-12-16

APPROVAL SIGNATURE: NPATEL P.M. Patel DATE: 12/12/2016 12-12-16