

CHROMOGENIX

Certificate of Analysis

PRODUCT:	CoaTest SP Factor FVIII Kit	PRODUCT # :	82408663
EXP. DATE:	2018-10	LOT # :	N1166470
VENDOR:	INSTRUMENTATION LABORATORY	ASSAYED PER QCM:	824086

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION
	MIN	MAX			
Coatest SP Factor VIII Kit (82103363)					
High Range					
Slope	0.0077	0.0112	--	0.0092	PASS
Intercept	--	0.250	--	0.161	PASS
Standard Curve (r)	0.99	--	--	0.994	PASS
Low Range					
Slope	0.020	0.060	--	0.039	PASS
Intercept	--	0.406	--	0.172	PASS
Standard Curve (r)	0.99	--	--	0.998	PASS

KIT COMPONENTS :			
NAME	LOT NUMBER	EXP. DATE	VOLUME
CaCl ₂	N1166215	2019-11	6 ml
CT SP Substrate	N1066060	2018-10	12 ml
Factor Reagent	N1066064	2019-04	10 ml
CT SP Buffer, pH 7.3	N1166252	2018-11	20 ml
CT SP Phospholipid	N1166468	2018-11	2 ml

PREPARED BY: NPATEL N.m. Patel DATE: 12/15/16 12-15-16

APPROVAL SIGNATURE: DTENCZA P. Dencza DATE: 12/15/16 12/15/16