

CHROMOGENIX

Certificate of Analysis

PRODUCT: CoaTest SP4 FVIII Kit	PRODUCT # : 82409463
EXP. DATE: 2018-08	LOT # : N0965322
VENDOR: INSTRUMENTATION LABORATORY	ASSAYED PER QCM: 824094

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION
	MIN	MAX			
Coatest SP4 Factor VIII Kit (82409463)					
High Range					
Slope	0.0077	0.0112	--	0.0096	PASS
Intercept	--	0.250	--	0.143	PASS
Standard Curve (r)	0.99	--	--	0.995	PASS
Low Range					
Slope	0.020	0.060	--	0.045	PASS
Intercept	--	0.406	--	0.236	PASS
Standard Curve (r)	0.99	--	--	0.996	PASS

KIT COMPONENTS :			
NAME	LOT NUMBER	EXP. DATE	VOLUME
CaCl ₂	N0864802	2019-08	(1x) 6 ml
CT SP Substrate	N0965020	2018-09	(1x) 12 ml
Factor Reagent	N0965157	2019-03	(4x) 3 ml
CT SP Buffer, pH 7.3	N0864795	2018-08	(1x) 20 ml
CT SP Phospholipid	N0864796	2018-08	(1x) 2 ml

PREPARED BY: NPATEL N.M. Patel DATE: 10/07/16 10-7-16

APPROVAL SIGNATURE: DTENCZA J. Dencza DATE: 10/07/16 10/7/16