

**CERTIFICATE OF ANALYSIS**

***CHROMOGENIX***

PRODUCT:	S-2302, 25 mg Substrate	PRODUCT # :	00082034039
EXP. DATE:	2019-10	LOT # :	N0663512
VENDOR:	INSTRUMENTATION LABORATORY	ASSAYED PER QCM:	820340

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION
	MIN	MAX			
SUBSTRATE CONCENTRATION	23.7	26.3	mg/vial	24.5	PASS
FREE p-NITROANILINE	0	0.5	mol %	0.0	PASS

DESCRIPTION	LOT NUMBER	EXP. DATE	VOLUME
S-2302	N0663512	2019-10	3 mL

PREPARED BY:

Maria Dittinger

DATE:

June 30, 2016

APPROVAL SIGNATURE:

J. Ameyra

DATE:

June 30, 2016