

CERTIFICATE OF ANALYSIS

CHROMOGENIX

PRODUCT:	S-2403, 25 mg Substrate	PRODUCT #:	82225439
EXP. DATE:	2018-08	LOT #:	N0858132
VENDOR:	INSTRUMENTATION LABORATORY	ASSAYED PER QCM:	822254

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION
	MIN	MAX			
SUBSTRATE CONCENTRATION	23.7	26.3	mg/vial	25.0	PASS
FREE p-NITROANILINE	0	0.5	mol %	0	PASS

DESCRIPTION	LOT NUMBER	EXP. DATE	VOLUME
S-2403	N0858132	2018-08	3 ml

PREPARED BY: *Maria Dettlinger* DATE: 9-8-15

APPROVAL SIGNATURE: *P. Longo* DATE: 9/10/15