

CERTIFICATE OF ANALYSIS

CHROMOGENIX

PRODUCT:	COATEST APC Resistance Control Plasma Level 2	PRODUCT # :	82266863
EXP. DATE:	2018-07	LOT # :	N0763871
VENDOR:	INSTRUMENTATION LABORATORY	ASSAYED PER QCM:	H01444

TEST DESCRIPTION	SPECIFICATIONS			RESULTS	DISPOSITION	
	MIN	MAX	UNITS			
Analysis with COATEST APC R:						
APT Time	CaCl ₂	27	44	seconds	43	PASS
	APC/CaCl ₂	--	101	seconds	84	PASS
APC Ratio		--	2.3	--	2.0	PASS

Analysis with COATEST APC R V:						
APT Time	CaCl ₂	27	40	seconds	38	PASS
APC Ratio		1.4	1.9	--	1.6	PASS

Kit Components:

DESCRIPTION	LOT NUMBER	EXP. DATE	VOLUME
Control Plasma Level 2	N0763871	2018-07	1 ml (x 5)

We hereby certify that the above product(s):

- have been manufactured and tested according to the requirements of their Manufacturing and Quality Control procedures and have been found to meet all conditions set forth with these documents;
- meet all applicable requirements of the 98/79/EC Directive according to which they are CE marked.

PREPARED BY: M. Oettinger *M. Oettinger*

DATE: 10/17/2016 10-17-16

APPROVAL SIGNATURE: D. Tencza *D. Tencza*

DATE: 10/17/2016 10/17/16