



Certificate of Analysis

PRODUCT:	CoaTest SP Factor FVIII Kit	PRODUCT # :	82408663
EXP. DATE:	2018-08	LOT # :	N0965029
VENDOR:	INSTRUMENTATION LABORATORY	ASSAYED PER QCM:	824086

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION
	MIN	MAX			
Coatest SP Factor VIII Kit (82103363)					
High Range					
Slope	0.0077	0.0112	--	0.0093	PASS
Intercept	--	0.250	--	0.138	PASS
Standard Curve (r)	0.99	--	--	0.998	PASS
Low Range					
Slope	0.020	0.060	--	0.044	PASS
Intercept	--	0.406	--	0.220	PASS
Standard Curve (r)	0.99	--	--	0.997	PASS

KIT COMPONENTS :			
NAME	LOT NUMBER	EXP.DATE	VOLUME
CaCl ₂	N0864802	2019-08	6 ml
CT SP Substrate	N0965020	2018-09	12 ml
Factor Reagent	N0965028	2019-03	10 ml
CT SP Buffer, pH 7.3	N0864795	2018-08	20 ml
CT SP Phospholipid	N0864796	2018-08	2 ml

PREPARED BY: MPATEL

DATE: 09/27/16

APPROVAL SIGNATURE: DTENCZA

DATE: 09/27/16