

CERTIFICATE OF ANALYSIS



DiaPharma Group, Inc.
8948 Beckett Road, West Chester, OH 45069

Product:	DiaPharma Factor X kit
Item Number:	DPGFX
Vendors:	R ² Diagnostics, Chromogenix
Lot #:	DPGFX049
Exp. Date:	2018-08-31

Kit Components:			
Name	Lot Number	Exp. Date	Volume (1 vial each)
FXa Chromogenic Substrate	N0965442	2019-09-30	25 mg
Russell's Viper Venom	DR-991	2018-08-31	<1 mg
CaCl ₂	DC-991	2018-08-31	20 ml
Tris Buffer	DB-991	2018-08-31	100 ml

Test Description	Specifications		Results		Disposition
	Min	Max	Units	Value	
Slope	1.0e-3 to 4.0e-3		NA	3.027e-03	Pass
Intercept	-.09 to .09		NA	3.108e-03	Pass
Regression Coefficient	R > 0.990		NA	0.999	Pass

Prepared By: *Monica A. Hurd* Date: 2017-03-13

Approved By: *Erin Reeves* Date: 2017-03-14