CERTIFICATE OF ANALYSIS

CHROMOGENIX

PRODUCT:	S-2222, 25 mg Substrate	PRODUCT # :	82031639
EXP. DATE:	2019-12	LOT # :	N0864682
VENDOR:	INSTRUMENTATION LABORATORY	ASSAYED PER QCM:	820316

TEST DESCRIPTION	SPECIFICATIONS			RESULTS	DISPOSITION
	MIN	MAX	UNITS		PASS/FAIL
SUBSTRATE CONCENTRATION	23.7	26.3	mg/vial	24.7	PASS
FREE p-NITROANILINE	0	0.5	mol %	0	PASS

DESCRIPTION	LOT NUMBER	EXP.DATE	VOLUME
S-2222	N0864682	2019-12	6 ml

PREPARED BY: *M. Oettinger* DATE: Aug 30,2016

APPROVAL SIGNATURE: *D. Tenczа* DATE: Aug 30,2016

Form QCM 820316-2 Rev. 01/09

526 Route 303, Orangeburg, New York 10962, Telephone 845-365-8000