

**CERTIFICATE OF ANALYSIS**

***CHROMOGENIX***

PRODUCT: COAMATIC FACTOR VIII	PRODUCT # : 82258563
EXP. DATE: 2018-12	LOT # : N0177165
VENDOR: INSTRUMENTATION LABORATORY	ASSAYED PER QCM: 822585/49730503
	DATE OF MANUFACTURE: 12/20/2016

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	PASS/FAIL
	MIN	MAX			
HIGH RANGE					
Slope	0.33	0.57	--	0.47	PASS
Intercept	--	0.12	--	0.09	PASS
r	0.990	--	--	0.999	PASS
LOW RANGE					
Slope	2.8	6.9	--	5.0	PASS
Intercept	--	0.21	--	0.15	PASS
r	0.990	--	--	0.999	PASS

**Kit Components:**

Name	Lot Number	Exp. date
S-2765/ I-2581	N0177163	2019-01
Factor Reagent	N0177164	2019-01
Buffer	N1266891	2018-12

We hereby certify that the above product(s):

- have been manufactured and tested according to the requirements of their Manufacturing and Quality Control procedures and have been found to meet all conditions set forth with these documents;
- meet all applicable requirements of the 98/79/EC Directive according to which they are CE marked.

PREPARED BY: MPATEL DATE: 1/23/2017

APPROVAL SIGNATURE: DTENCZA DATE: 1/23/2017