

CERTIFICATE OF ANALYSIS

CHROMOGENIX

PRODUCT: S-2765, 25 mg Substrate	PRODUCT #: 82141339
EXP. DATE: 2019-09	LOT #: N0965442
VENDOR: INSTRUMENTATION LABORATORY	ASSAYED PER QCM: 821413

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION
	MIN	MAX			
SUBSTRATE CONCENTRATION	23.7	26.3	mg/vial	24.8	PASS
FREE p-NITROANILINE	0	0.5	mol %	0	PASS

DESCRIPTION	LOT NUMBER	EXP. DATE	VOLUME
S-2765	N0965442	2019-09	3 ml

PREPARED BY: Maria Ottengjer

DATE: Oct 14, 2016

APPROVAL SIGNATURE: J. Amey

DATE: 10/14/16