

CERTIFICATE OF ANALYSIS

CHROMOGENIX

PRODUCT:	S-2238, 25 mg Substrate	PRODUCT #:	82032439
EXP. DATE:	2020-01	LOT #:	N0177413
VENDOR:	INSTRUMENTATION LABORATORY	ASSAYED PER QCM:	820324

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION
	MIN	MAX			
SUBSTRATE CONCENTRATION	23.7	26.3	mg/vial	25.3	PASS
FREE p-NITROANILINE	0	0.5	mol %	0.1	PASS

DESCRIPTION	LOT NUMBER	EXP. DATE	VOLUME
S-2238	N0177413	2020-01	6 ml

PREPARED BY: Maia Dettner DATE: 2-7-17

APPROVAL SIGNATURE: P. Encya DATE: 2/8/17