

CERTIFICATE OF ANALYSIS

CHROMOGENIX

PRODUCT: COATEST APC Resistance Control Plasma Level 1 PRODUCT #: 82265063
 EXP. DATE: 2019-06 LOT #: N0570180
 VENDOR: INSTRUMENTATION LABORATORY ASSAYED PER QCM: C00423

TEST DESCRIPTION	SPECIFICATIONS		RESULTS	DISPOSITION
	MIN	MAX		
Analysis with COATEST APC R: APC Time CaCl ₂	28	34	30	PASS
APC/CaCl ₂	90	--	116	PASS
APC Ratio	3.5	4.4	3.8	PASS

Analysis with COATEST APC R V: APC Time CaCl ₂	--	38	34	PASS
APC/CaCl ₂	75	--	105	PASS
APC Ratio	2.7	3.2	3.1	PASS

Kit Components:

DESCRIPTION	LOT NUMBER	EXP. DATE	VOLUME
Control Plasma Level 1	N0570180	2019-06	1 ml (x 5)

We hereby certify that the above product(s):

- have been manufactured and tested according to the requirements of their Manufacturing and Quality Control procedures and have been found to meet all conditions set forth with these documents;
- meet all applicable requirements of the 98/79/EC Directive according to which they are CE marked.

PREPARED BY: N. Patel N.M. Patel DATE: 6/15/2017 6:15:17

APPROVAL SIGNATURE: D. Tenenza DATE: 6/15/2017 6/15/17