

**CERTIFICATE OF ANALYSIS**

**CHROMOGENIX**

**Instrumentation  
Laboratory**

A Werfen Company

PRODUCT NAME:	<u>S-2251, 25 mg Substrate</u>	EXPIRATION DATE:	<u>2024-04</u>
PRODUCT NUMBER:	<u>82033239</u>		
LOT:	<u>N0412235</u>		

TEST DESCRIPTION	SPECIFICATIONS			RESULTS	DISPOSITION
	MIN	MAX	UNITS		
SUBSTRATE CONCENTRATION	23.7	26.3	mg/vial	25.1	PASS
FREE p-NITROANILINE	0	0.5	mol %	0	PASS

PREPARED BY: Maria Oettinger *Maria Oettinger*  
 PRINT SIGNATURE

DATE: 5-17-21

TITLE: QC Analyst III

APPROVAL SIGNATURE: Debra Harz *Debra Harz*  
 PRINT SIGNATURE

DATE: 5-17-21

TITLE: QC Technical Supervisor