

CERTIFICATE OF ANALYSIS

CHROMOGENIX


**Instrumentation
Laboratory**
 A Werfen Company

PRODUCT NAME:	<u>Antithrombin 1 x 25 IU</u>	EXPIRATION DATE:	<u>2026-08-31</u>
PRODUCT NUMBER:	<u>81079639</u>		
LOT NUMBER:	<u>N0844864</u>		

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION
	MIN	MAX			
Antithrombin Activity	25	30	IU/vial	29	PASS
Assay Slope @ 37°C	-0.0024	-0.0034		-0.0028	PASS
Intercept @ 37°C	0.40	0.60		0.47	PASS
r ²	0.990	--		0.999	PASS

PREPARED BY: Monika Skowronska Monika Skowronska
 PRINT SIGNATURE

DATE: 8/28/2024 8/28/24

TITLE: QC Analyst II

APPROVAL SIGNATURE: Nehal Patel N.M. Patel
 PRINT SIGNATURE

DATE: 8/28/2024 8-28-24

TITLE: Senior Group Leader