



**ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS**

DQRH00AZ

TECHNOZYM® D-Dimer ELISA Kit 96T

REF 2599006
 LOT RH61B00.01
 ⌚ 30-Sep-2017

Inhalt Content	Reagenz Reagent	LOT	⌚
1 x	D-Dimer Plate	RKS52B0	30-Sep-2017
1x0.3mL	D-Dimer Conjugate	RKV41B1	31-Dec-2017
1x80mL	Wash buffer Concentrate	RNA52B0	31-Oct-2017
1x90mL	Incubation buffer	RNB61B0	31-Jan-2018
1x12mL	TMB Substrate	RKN52B0	31-Dec-2017
1x12mL	Stop solution	RKK52B0	31-Dec-2017
1x0.5mL	D-Dimer Calibrator 1	1KT31B2	31-Dec-2017
1x0.5mL	D-Dimer Calibrator 2	2KT31B2	31-Dec-2017
1x0.5mL	D-Dimer Calibrator 3	3KT31B2	31-Dec-2017
1x0.5mL	D-Dimer Calibrator 4	4KT31B2	31-Dec-2017
1x0.5mL	D-Dimer Calibrator 5	5KT42B1	31-Dec-2017
1x0.5mL	D-Dimer High Control	RKY31B2	31-Dec-2017
1x0.5mL	D-Dimer Low Control	RKZ31B2	31-Dec-2017

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	> 1.0	1.1	OK
	Calibrator 5	< 0.1	0.04	OK
Verwendbarkeit Usability	High Control (RKY31B2)	881.4 – 1469.0 ng/mL	1251.93	OK
	Low Control (RKZ31B2)	208.0 – 346.7 ng/mL	338.29	OK

Durchführung von
Prepared by

Fikret NASUFI *Fikret Nasufi*

18-Feb-2016
Datum / Date:

Freigabe von
Approved by

Christoph ENNE *Christoph Enne*

18-Feb-2016
Datum / Date: