
	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ6T0AAZ.02
---	---	-------------

Ceveron Technoplastin HIS 12x4mL

REF 5003010
LOT 6T53B.01
 31-May-2017

Analytische Verfahren Analytical Procedure			Anforderung Requirement	Ergebnis Result	
Heparinempfindlichkeit Heparin Sensitivity	Abweichung zur 0 E/mL UFH Deviation to 0 E/mL UFH	1 E/mL UFH	± 10%	- 2.68	OK
Verwendbarkeit Technoplastin HIS Usability Technoplastin HIS Ceveron® alpha	Technoplastin HIS %	CCN LOT 1P37B00	109.7 – 138.3%	117.5	OK
	Technoplastin HIS INR	CCN LOT 1P37B00	0.77 – 0.97	0.85	OK
	Technoplastin HIS %	CCA LOT 3P38B00	25.7 – 32.3%	28.1	OK
	Technoplastin HIS INR	CCA LOT 3P38B00	2.54 – 3.20	2.98	OK
Etiketten und Barcode Überprüfung Label and bar code Verification				OK	

**Durchführung von
Prepared by**

Fikret NASUFI

Fikret Nasufi

01-Jun-2015
Datum / Date:

**Freigabe von
Approved by**

Lavdrim NASUFI

Lavdrim Nasufi

01-Jun-2015
Datum / Date: