

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ6C0A0AZ.03
---	---	--------------

TECHNOCLOT® PT PLUS 5x4ml

REF 5003220
LOT 6C46BA0.01
 31-Jul-2016

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Kontrollwiederfindung Recovery of controls	CCN LOT 1P44B00	85.0 – 107% INR: 0.90 – 1.14	88.2 1.06	OK
	CCA LOT 2B35B00	14.6 – 18.4% INR: 2.41 – 3.03	16.3 2.84	OK
Verwendbarkeit Usability Ceveron® alpha	CNN LOT 1P44B00	0.90 – 1.14 INR	0.95	OK
	CCA LOT 3P38B00	2.47 – 3.11 INR	2.67	OK
Etiketten und Barcode Überprüfung Label and bar code verification			OK	

Durchführung von
Prepared by


Lucas HOLOVICZ

18-Sep-2014
Datum / Date:

Freigabe von
Approved by


Fikret NASUFI

18-Sep-2014
Datum / Date: