

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ7A00AZ
---	---	----------

TECHNOTHROMBIN[®] TGA RA Reagenz 5x0.5mL

REF 5006205
LOT 7A23B01.01
 31-Jul-2016

Inhalt Content	Reagenz Reagent	LOT	
5x0.5 mL	TGA RA Reagenz	7A23B01	31-Jul-2016

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Kontrollwiederfindung Recovery of controls	TGA Control High (nM Peak Thrombin) LOT 1B41B00	254.8 – 424.6	258.3	OK

Durchführung von
Prepared by

Fikret NASUFI


15-Jul-2014

Datum / Date:

Freigabe von
Approved by

Michael RIHA


15-Jul-2014

Datum / Date: