
	<p style="text-align: center;">ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS</p>	<p style="text-align: center;">DQ7C00AZ</p>
---	---	---

TECHNOTHROMBIN[®] TGA RB 5x0.5mL

REF 5006209
LOT 7C31B00.01
 31 - Jul - 2015

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.0 – 7.6	7.25	OK
Restfeuchte Residual moisture		$\leq 1 \%$	0.090	OK
Homogenität Homogeneity	TGA Control High (nM Peak Thrombin) LOT 1B12B01)	CV $\leq 10 \%$	6.0	OK
Kontrollwiederfindung Recovery of controls	TGA Control High (nM Peak Thrombin) LOT 1B12B01)	186 - 346	326.12	OK

Durchführung von
Prepared by

Fikret NASUFI

F. Nasufi

26-Sep-2013
Datum / Date:

Freigabe von
Approved by

Nikolaus BINDER

Nikolaus Binder

26. SEP. 2013
Datum / Date: