
	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ7C00AZ
---	---	----------

TECHNOTHROMBIN® TGA RB 5x0.5mL

REF 5006209

LOT 7C51B00.01

 31-Jul-2017

Inhalt Content	Reagenz Reagent	LOT	
5x0.5 mL	TGA RB Reagenz	7C51B00	31-Jul-2017


Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result	
Kontrollwiederfindung Recovery of controls	254 - 472	421.7	OK

Durchführung von
Prepared by



11-Aug-2015
Datum / Date:

Freigabe von
Approved by



11-Aug-2015
Datum / Date: