
	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ7C00AZ LOT 7C91A01
---	---	------------------------------------

TECHNOTHROMBIN[®] TGA RB 50x0.5 mL

REF 5006210
LOT 7C91A01.01
 30 - Apr - 2015

Analytische Verfahren Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability	TGA Control High (nM Peak Thrombin) LOT 1B12B01	316 – 586	467.5	OK

Durchführung von
Prepared by

F. Nasubi

07-Mai-2013
Datum / Date:

Freigabe von
Approved by

Harald SEVCIK
H. Sevcik

08. MAI 2013
Datum / Date: