

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ7D00AZ
---	---	----------

TECHNOTHROMBIN[®] TGA RC Low 5x0.5mL

REF 5006212

LOT 7D31B00.01

 31 - Mar - 2015

Inhalt Content	Reagenz Reagent	LOT	
5x0.5 mL	TGA RC Low	7D31B00	31-Mar-2015

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Kontrollwiederfindung Recovery of controls	TGA Control High (nM Peak Thrombin) LOT 1B12B01	231 - 429	305.6	OK
	TGA Control Low (nM Peak Thrombin) LOT 2B34B00	69 - 128	79.1	OK

Durchführung von
Prepared by

Fikret NASUFI

F. Nasufi

15-Oct-2013
Datum / Date:

Freigabe von
Approved by

Michael RIHA

Michael Riha

15-Oct-2013
Datum / Date: