

	<b>ANALYSENZERTIFIKAT</b> <b>CERTIFICATE OF ANALYSIS</b>	DQ7D00AZ
---	---	----------

## TECHNOTHROMBIN® TGA RC Low 50x0.5mL

REF      5006213  
LOT      7D41B00.01  
      31 - Mai - 2016

<b>Inhalt</b> <b>Content</b>	<b>Reagenz</b> <b>Reagent</b>	<span style="border: 1px solid black; padding: 2px;">LOT</span>	
50x0.5 mL	TGA RC Low	7D41B00	31-Mai-2016

<b>Analytische Prozedur</b> <b>Analytical Procedure</b>		<b>Anforderung</b> <b>Requirement</b>	<b>Ergebnis</b> <b>Result</b>	
Kontrollwiederfindung Recovery of controls	TGA Control High (nM Peak Thrombin) <span style="border: 1px solid black; padding: 2px;">LOT</span> 1B41B00	276 - 512	325.2	OK

**Durchführung von**  
**Prepared by**

**Fikret NASUFI**   


---

03-Dec-2014  
**Datum / Date:**

**Freigabe von**  
**Approved by**

**Michael RIHA**   


---

03-Dec-2014  
**Datum / Date:**