

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ7D00AZ
---	---	----------


TECHNOTHROMBIN[®] TGA RC Low 50x0.5mL

REF 5006213
LOT 7D51B00.02
 28-Feb-2017

Inhalt Content	Reagenz Reagent	LOT	
50x0.5 mL	TGA RC Low	7D51B00	28-Feb-2017

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Kontrollwiederfindung Recovery of controls	TGA Control High (nM Peak Thrombin) LOT 1B41B00	190 - 352	222.1	OK

**Durchführung von
Prepared by**

Fikret NASUFI


02-Dec-2015

Datum / Date:

**Freigabe von
Approved by**

Michael KAFKA


02-DEC-2015

Datum / Date: