



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQ7F00AZ

TECHNOTHROMBIN[®] TGA RC High 5x0.5mL

REF 5006214

LOT 7F31B00.01

⌚ 31 - Mar - 2015

Inhalt Content	Reagenz Reagent	LOT	⌚
5x0.5 mL	TGA RC High	7F31B00	31-Mar-2015

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Kontrollwiederfindung Recovery of controls	TGA Control High (nM Peak Thrombin) LOT 1B12B01)	316 - 586	433.1	OK
	TGA Control Low (nM Peak Thrombin) LOT 2B34B00	132 - 245	197.5	OK

Durchführung von
Prepared by

Fikret NASUFI

F. Nasufi

18-Oct-2013
Datum / Date:

Freigabe von
Approved by

Lavdrim NASUFI

Lavdrim Nasufi

18-Oct-2013
Datum / Date: