

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ7F00AZ
---	---	----------

TECHNOTHROMBIN® TGA RC High 5x0.5mL

REF 5006214
LOT 7F42B00.01
 29 - Feb - 2016

Inhalt Content	Reagenz Reagent	LOT	
5x0.5 mL	TGA RC High	7F42B00	29-Feb-2016

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Kontrollwiederfindung Recovery of controls	TGA Control High (nM Peak Thrombin) LOT 1B41B00	276 - 512	477.7	OK
	TGA Control Low (nM Peak Thrombin) LOT 2B43B00	137 - 255	145.7	OK

**Durchführung von
Prepared by**

Fikret NASUFI 

17-Jul-2014
Datum / Date:

**Freigabe von
Approved by**

Michael RIHA 

17-Jul-2014
Datum / Date: