

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ7F00AZ
---	---	----------

TECHNOTHROMBIN® TGA RC High 5x0.5mL

REF 5006214
LOT 7F51B00.01
 30-Apr-2017

Inhalt Content	Reagenz Reagent	LOT	
5x0.5 mL	TGA RC High	7F51B00	30-Apr-2017

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Kontrollwiederfindung Recovery of controls	TGA Control High (nM Peak Thrombin) LOT 1B41B00)	276 - 512	370.6	OK


Durchführung von
Prepared by

Fikret NASUFI *Fikret Nasufi*

28-Apr-2015

Datum / Date:

Freigabe von
Approved by


Lavdrim NASUFI

28-Apr-2015

Datum / Date: