

	<b>ANALYSENZERTIFIKAT</b> <b>CERTIFICATE OF ANALYSIS</b>	DQ7F00AZ
---	---	----------

## TECHNOTHROMBIN® TGA RC High 50x0.5mL

REF      5006216  
LOT      7F31B00.01  
      31 - Mar - 2015

<b>Inhalt</b> <b>Content</b>	<b>Reagenz</b> <b>Reagent</b>	<span style="border: 1px solid black; padding: 2px;">LOT</span>	
50x0.5 mL	TGA RC High	7F31B00	31-Mar-2015

<b>Analytische Prozedur</b> <b>Analytical Procedure</b>		<b>Anforderung</b> <b>Requirement</b>	<b>Ergebnis</b> <b>Result</b>	
Kontrollwiederfindung Recovery of controls	TGA Control High (nM Peak Thrombin) <span style="border: 1px solid black; padding: 2px;">LOT</span> 1B12B01)	316 - 586	483.5	OK
	TGA Control Low (nM Peak Thrombin) <span style="border: 1px solid black; padding: 2px;">LOT</span> 2B02BRV	55 - 101	90.6	OK

**Durchführung von**  
**Prepared by**

*F. Nasari*

*14-Jun-2013*  
Datum / Date:

**Freigabe von**  
**Approved by**

Harald SEVCIK

*14. JUNI 2013*  
Datum / Date: