

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ7G00AZ
---	---	----------

TECHNOTHROMBIN® TGA Reagent RD 5x1.5mL

REF 5006220
LOT 7G31B00.01
 30 - Jun - 2015

Inhalt Content	Reagenz Reagent	LOT	
5x1.5mL	TGA Reagent RD	7G31B00	30-Jun-2015

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Kontrollwiederfindung Recovery of controls	TGA Control High (nM Peak Thrombin) LOT 1B41B00)	161 - 299	273.9	OK
	TGA Control Low (nM Peak Thrombin) LOT 2B43B00	239 - 443	361.2	OK

**Durchführung von
Prepared by**

Fikret NASUEF 

01-Sep-2014
Datum / Date:

**Freigabe von
Approved by**

Michael RIHA 

01-SEP-2014
Datum / Date: