

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ7G00AZ
---	---	----------

TECHNOTHROMBIN® TGA Reagent RD 50x1.5mL

REF 5006222
LOT 7G32B00.01
 31 - Dec - 2015

Inhalt Content	Reagenz Reagent	LOT	
50x1.5mL	TGA Reagent RD	7G32B00	30-Dec-2015

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Kontrollwiederfindung Recovery of controls	TGA Control High (nM Peak Thrombin) LOT 1B12B01)	166 - 308	259.5	OK
	TGA Control Low (nM Peak Thrombin) LOT 2B34B00	202 - 376	271.4	OK

**Durchführung von
Prepared by**

Fikret NASUFI

Fikret Nasufi

13-Feb-2014
Datum / Date:

**Freigabe von
Approved by**

Michael RIHA

Michael Riha

13-FEB-2014
Datum / Date: