

	<b>ANALYSENZERTIFIKAT</b> <b>CERTIFICATE OF ANALYSIS</b>	DQ7G00AZ
---	---	----------

## TECHNOTHROMBIN<sup>®</sup> TGA RD 5x1.5mL

REF      5006220  
LOT      7G51B00.01  
      31-Jan-2017

Inhalt Content	Reagenz Reagent	<span style="border: 1px solid black; padding: 2px;">LOT</span>	
5x1.5mL	TGA Reagent RD	7G51B00	31-Jan-2017

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result	
Kontrollwiederfindung Recovery of controls	TGA Control High (nM Peak Thrombin) <span style="border: 1px solid black; padding: 2px;">LOT</span> 1B41B00 161 - 299	260.93	OK
	TGA Control Low (nM Peak Thrombin) <span style="border: 1px solid black; padding: 2px;">LOT</span> 2B43B00 239 - 443	389.63	OK

**Durchführung von**  
**Prepared by**

*Fikret NASUFI*

*16-Jun-2015*  
Datum / Date:

**Freigabe von**  
**Approved by**

*[Signature]*

*16-Jun-2015*  
Datum / Date: