

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ7D00AZ
---	---	----------

Ceveron[®] TGA Reagenz RC Low 5x1 mL

REF 5006224
LOT 7D42B.01
 30 - Jun - 2016

Inhalt Content	Reagenz Reagent	LOT	
5x1 mL	Ceveron [®] TGA RC Low	7D42B	30-Jun-2016

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result	
Kontrollwiederfindung Recovery of controls	TGA Control High (nM Peak Thrombin) LOT 1B41B00	106 - 198	183.6 OK
	TGA Control High (Lag time min) LOT 1B41B00	3.2 - 6.0	3.8 OK
	TGA Control High (AUC) LOT 1B41B00	2001 - 3715	2456.3 OK
Etiketten und Barcode Überprüfung Label and bar code Verification		OK	

Durchführung von
Prepared by

Fikret NASUFI

Fikret Nasufi

30-Jul-2014
Datum / Date:

Freigabe von
Approved by

Michael RIHA

Michael Riha

30-Jul-2014
Datum / Date: