

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ7F00AZ
---	---	----------

Ceveron® TGA RC High 5x1 mL

REF 5006227
LOT 7F52B.01
 31-Aug -2017

Inhalt Content	Reagenz Reagent	LOT	
5x1 mL	Ceveron® TGA RC High	7F52B	31-Aug-2017

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result		
Kontrollwiederfindung Recovery of controls	TGA Control High (nM Peak Thrombin) LOT (1B41B00)	188 – 350	192.9	OK
	TGA Control High (AUC) LOT (1B41B00)	2082 – 3866	2209.5	OK
	TGA Control Low (nM Peak Thrombin) LOT (2B43B00)	47 - 87	57.5	OK
	TGA Control Low (AUC) LOT (2B43B00)	527 - 979	650.6	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von
Prepared by

Fikret NASUFI

12. NOV - 2015
Datum / Date:

Freigabe von
Approved by

Michael KAFKA *[Signature]*

12. NOV - 2015
Datum / Date: