




ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS


DQ8F00AZ

TECHNOTHROMBIN® TGA SUB 50x1.5mL

REF 5006230

LOT 8F31B00.02

 28 - Feb - 2015

Inhalt Content	Reagenz Reagent	LOT	
50x1.5mL	TGA Substrat	8F31B00	28-Feb-2015

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Kontrollwiederfindung RC High Recovery of controls RC High	TGA Control High (nM Peak Thrombin) LOT 1B12B01)	316.0 – 586.0	503.2	OK
Kontrollwiederfindung RC Low Recovery of controls RC Low	TGA Control High (nM Peak Thrombin) LOT 1B12B01)	231.0 - 429	405.2	OK
Kontrollwiederfindung RD Recovery of controls RD	TGA Control High (nM Peak Thrombin) LOT 1B12B01)	166.0 – 308.0	236.6	OK

Durchführung von
Prepared by

Fikret NASUFI




20-Feb-2014

Datum / Date:

Freigabe von
Approved by

Michael RIHA



20-FEB-2014

Datum / Date: