

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ8F00AZ
---	---	----------

TECHNOTHROMBIN® TGA SUB 50x1.5mL

REF 5006230
LOT 8F31B01.01
 30-Apr-2016

Inhalt Content	Reagenz Reagent	LOT	
50x1.5mL	TGA Substrat	8F31B01	30-Apr-2016

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result		
Kontrollwiederfindung Recovery of controls	TGA Control High (nM Peak Thrombin) LOT 1B41B00	TGA RC High LOT 7F42B00	276 - 512	449.1	OK
	TGA Control Low (nM Peak Thrombin) LOT 2B43B00	TGA RC High LOT 7F42B00	137 - 255	140.8	OK

Durchführung von
Prepared by

Fikret NASUFI


02 - Oct - 2014
Datum / Date:

Freigabe von
Approved by

Michael RIHA


02 - OCT - 2014
Datum / Date: