



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQ8F00AZ

TECHNOTHROMBIN® TGA SUB 50x1.5mL

REF 5006230

LOT 8F41B00.03

⌚ 31-Mai-2016

Inhalt Content	Reagenz Reagent	LOT	⌚
50x1.5mL	TGA Substrat	8F41B00	31-Mai-2016

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result		
Kontrollwiederfindung Recovery of controls	TGA Control High (nM Peak Thrombin) LOT 1B41B00	TGA RC High LOT 7F42B00	276 - 512	465.2	OK
		TGA RC Low LOT 7D41B00	190 - 352	322.9	OK
		TGA RB LOT 7C31B00	254 - 472	333.5	OK
		TGA RD LOT 7G31B00	161 - 299	208.1	OK
	TGA Control Low (nM Peak Thrombin) LOT 2B43B00	TGA RC High LOT 7F42B00	137 - 255	214.1	OK
		TGA RC Low LOT 7D41B00	48 - 88	56.2	OK
		TGA RD LOT 7G31B00	239 - 443	298.9	OK

Durchführung von
Prepared by

Fikret NASUFI
Fikret Nasufi

13-Oct-2014
Datum / Date:

Freigabe von
Approved by

Michael RIHA
Michael Riha

15-Oct-2014
Datum / Date: