
	<b>ANALYSENZERTIFIKAT</b> <b>CERTIFICATE OF ANALYSIS</b>	DQ8F00AZ
---	---	----------

## TECHNOTHROMBIN<sup>®</sup> TGA SUB 50x1.5mL

REF      5006230  
LOT      8F52B00.01  
      31-Oct-2017

Inhalt Content	Reagenz Reagent	<span style="border: 1px solid black; padding: 2px;">LOT</span>	
50x1.5mL	TGA Substrat	8F52B00	31-Oct-2017

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result		
Kontrollwiederfindung Recovery of controls	TGA Control High (nM Peak Thrombin)	TGA RC High <span style="border: 1px solid black; padding: 2px;">LOT</span> 7F51B00	276 - 512	317.7	OK
	<span style="border: 1px solid black; padding: 2px;">LOT</span> 1B41B00	TGA RC Low <span style="border: 1px solid black; padding: 2px;">LOT</span> 7D51B00	190 - 352	220.5	OK

**Durchführung von**  
**Prepared by**

Fikret NASUFI  
*Fikret Nasufi*

27-Nov-2015  
Datum / Date:

**Freigabe von**  
**Approved by**

Michael KAFKA  
*Michael Kafka*

27-Nov-2015  
Datum / Date: