

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ8F00AZ
---	---	----------

TECHNOTHROMBIN[®] TGA SUB 5x1.5mL

REF 5006235

LOT 8F31B00.01

 28 - Feb - 2015

Inhalt Content	Reagenz Reagent	LOT	
5x1.5mL	TGA Substrat	8F31B00	28-Feb-2015

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result	
Kontrollwiederfindung RC High	TGA Control High (nM Peak Thrombin) LOT 1B12B01)	316.0 – 586.0	454.7 OK
Recovery of controls RC High	TGA Control Low (nM Peak Thrombin) LOT 2B34B00	132.4 – 245.8	193.8 OK
Kontrollwiederfindung RC Low	TGA Control High (nM Peak Thrombin) LOT 1B12B01)	231.0 - 429	286.0 OK
Recovery of controls RC Low	TGA Control Low (nM Peak Thrombin) LOT 2B34B00	69.4 – 128.7	71.8 OK
Kontrollwiederfindung RD	TGA Control High (nM Peak Thrombin) LOT 1B12B01)	166.0 – 308.0	286.1 OK
Recovery of controls RD	TGA Control Low (nM Peak Thrombin) LOT 2B34B00	202.8 – 376.7	271.2 OK

Durchführung von
Prepared by

Fikret NASUFI

F. Nasufi

10-Sep-2013

Datum / Date:

Freigabe von
Approved by

Nikolaus BINDER

Nikolaus Binder

10. SEP. 2013

Datum / Date: